



Member State  
Denmark

**OIML Certificate N°  
R76/2006-DK3-12.04**

## **OIML CERTIFICATE OF CONFORMITY**

### **Issuing authority**

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgård Jensen

### **Applicant**

Name: **Kingship Weighing Machine Corp.**  
Address: 739, Renhua Rd.  
Dali Dist., Taichung City 412  
Taiwan R.O.C.

### **Manufacturer**

Name: **Kingship Weighing Machine Corp.**  
Address: 739, Renhua Rd.  
Dali Dist., Taichung City 412  
Taiwan R.O.C.

### **Identification**

of the certified pattern: **Non-automatic weighing instrument**  
**Type: GEC / GEW / KSP / KSPL / MEP / MEPL**  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 2006**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages



## OIML Certificate N° R76/2006-DK3-12.04

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1912657, dated 3<sup>rd</sup> Dec. 2012, including 88 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
21 December 2012

J. Hovgård Jensen  
Certification Officer

### Characteristics

Accuracy class		III
Weighing range		Single-interval or multiple range or multi-interval (2 ranges/intervals)
Maximum capacity	Max	3 kg to 30 kg
Minimum capacity	Min	20 e
Verification scale interval	e	$\geq 1$ g
Number of intervals (per partial weighing range/interval)	$\text{Max}_i / e_i$	$\leq 3000$
Tare	T	$\leq -\text{Max}$
Temperature range		-10 °C to +40 °C

### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.