



Member State
Denmark

OIML Certificate N°
R76/2006-DK3-14.05 revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**
Address: Venlighedsvej 4
2970 Hørsholm
Denmark

Person responsible: J. Hovgård Jensen

Applicant

Name: **OHAUS Corporation**
Address: 7 Campus Drive, Suite 310
Parsippany, NJ 07054
USA

Manufacturer

of the certified pattern: **OHAUS Corporation**

Identification

of the certified pattern: **Non-automatic weighing instrument**
Type: A31P
Intended to be used for direct sales to the public.
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 2006
for accuracy class III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages



OIML Certificate N° R76/2006-DK3-14.05 revision 1

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1912639, dated 26-11-2012 that includes 76 pages, DELTA, DK, No. DANAK-1913414, dated 14-10-2013 that includes 20 pages, DELTA, DK, No. DANAK-1915400, dated 25-06-2015 that includes 20 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**
17 August 2015

J. Hovgård Jensen
Certification Officer

Characteristics

Accuracy class		III
Weighing range		Single-interval or dual-range or dual-interval
Maximum capacity	Max	3 kg to 30 kg
Minimum capacity	Min	20 e
Verification scale interval	e	≥ 0.5 g
Number of intervals	Max_i / e_i	≤ 3000
Subtractive tare	T	$\leq -\text{Max}$
Interface		Protective, according to paragraph 5.3.6
Supply voltage		100-240 VAC, 50/60 Hz using external AC/DC adapter with 12 VDC output, optional 6 V internal battery
Temperature range for the indicator		-10 °C / +40 °C
Software identification:		Ver 1.01 or Ver 1.02

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.