



Member State  
Denmark

OIML Certificate N°  
R76/2006-DK3-15.03

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgård Jensen

### Applicant

Name: **CAS Corporation**  
Address: #262, Geurugogae-ro, Gwangjeok-myeon,  
Yangju-si, Gyeonggi-do  
REPUBLIC OF KOREA

### Manufacturer

of the certified pattern: **CAS (Zhejiang) Electronics Co., Ltd.**  
99# Changjiang Road, Jiashan County, Zhejiang Province,  
CHINA

### Identification

of the certified pattern: **Non-automatic weighing instrument**  
**Type: XE**  
Intended to be used for direct sales to the public.  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 2006**  
**for accuracy class II**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

DELTA



## OIML Certificate N° R76/2006-DK3-15.03

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1914018, dated 04-04-2014 that includes 77 pages, DELTA, DK, No. DANAK-1914528, dated 30-09-2014 that includes 72 pages, DELTA, DK, No. DANAK-1914991, dated 12-02-2015 that includes 41 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
26 May 2015

J. Hovgård Jensen  
Certification Officer

### Characteristics

Accuracy class		II
Weighing range		Single-interval
Maximum capacity	Max	300 g to 6000 g
Minimum capacity	Min	According to OIML R76-1:2006 table 3 and section 3.4.3
Verification scale interval	e	$\geq 0.05$ g
Display resolution	d	$0.1 \times e$ ( $\geq 0.005$ g) as auxiliary indicating device for model XE600 and XE6000, as extended resolution for remaining models
Number of intervals	Max / e	6000, or 7500 for XE 1500
Subtractive tare	T	$\leq -\text{Max}$
Interface		Protective, according to paragraph 5.3.6
Supply voltage		100-240 VAC/12 VDC, 50/60 Hz using external AC/DC adapter
Temperature range for the indicator		+5 °C / +40 °C
Software identification:	Version	2.00

#### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.