

Member State  
Denmark

**OIML Certificate N°  
R76/2006-DK3-16.07**

## **OIML CERTIFICATE OF CONFORMITY**

### **Issuing authority**

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgård Jensen

### **Applicant**

Name: **GSS SCALE (SUZHOU) CO., Ltd.**  
Address: No.1, Jinrui Road, Taiping Industrial Park,  
Xiangcheng District, Suzhou,  
CHINA

### **Manufacturer**

of the certified pattern: **GSS SCALE (SUZHOU) CO., Ltd.**

### **Identification**

of the certified pattern: **Non-automatic price computing weighing instrument**  
**Type: LWN / LWNT / AIPI-SS / AIPIT-SS / AIPI-SS2 / AIPIT-SS2 /**  
**LCN / LCNT / HWN / HWNT / LHW-SZ / LHWT-SZ**  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 2006**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

DELTA

# OIML Certificate N° R76/2006-DK3-16.07

The conformity was established by tests described in the associated test report, DELTA, DK, No. DANAK-1915955, dated 09-12-2015, 91 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
16 August 2016



J. Hovgård Jensen  
Certification Officer

## Characteristics

Accuracy class		III
Weighing range		Single-interval or dual-range
Maximum capacity	Max <sub>i</sub>	3 kg to 30 kg
Minimum capacity	Min <sub>i</sub>	20 e
Verification scale interval	e <sub>i</sub>	≥ 1 g
Number of intervals	Max <sub>i</sub> / e <sub>i</sub>	≤ 3000
Subtractive tare	T	≤ -Max
Interface		Protective, according to paragraph 5.3.6
Supply voltage		12 VDC via external adapter with input 100-240 VAC 50/60Hz, or 100-240VAC 50/60Hz. optional 6 V internal battery
Temperature range for the indicator		-10 °C / +40 °C
Software identification:	LCN/LCNT all other	VER-01 VER-11

### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.