



Member State
Denmark

OIML Certificate N°
R76/2006-DK3-17.05

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**
Address: Venlighedsvej 4
2970 Hørsholm
Denmark

Person responsible: J. Hovgård Jensen

Applicant

Name: **Shandong New Beiyang Information Technology Co., Ltd.**
Address: 169 Huoju Road, HDZ, Weihai,
Shandong 264209
CHINA

Manufacturer

of the certified pattern: **Shandong New Beiyang Information Technology Co., Ltd.**

Identification

of the certified pattern: **Non-automatic weighing instrument**
Type: BWM-S1000 / BWM-S1100 / BWM-S1200 / BWM-S2000 /
BWM-S3000 / BWM-S4000 / BWM-S5000 / BWM-S6000 /
BWM-S8000
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 2006
for accuracy class III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

DELTA



OIML Certificate N° R76/2006-DK3-17.05

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1913578, dated 01-11-2013 that includes 71 pages
DELTA, DK, No. DANAK-1914341, dated 16-07-2014 that includes 37 pages

The issuing authority: **DELTA, OIML Issuing Authority DK3**
14 June 2017

J. Hovgård Jensen
Certification Officer

Characteristics

| | | |
|---|------------------|--|
| Accuracy class | | III |
| Weighing range | | Single-interval |
| Maximum capacity | Max _i | 200 kg |
| Number of intervals | Max / e | 4000 |
| Verification scale interval | e = | 0.05 kg |
| Subtractive tare | T | ≤ -Max |
| Load receptor | | Platform w/wo rollers |
| Interface | | Protective, according to paragraph 5.3.6 |
| Connected load cells | | 4 pcs. ZEMIC H8C C5 200 kg (R60/2000-NL1-11.30) |
| Supply voltage | | 200-240 VAC, AC adapter, output 12 VDC 500 mA |
| Temperature range for the indicator | | -10 °C / +40 °C |
| Software identification: | | Legally relevant: 1.10 Application: 1.xx |
| Max length of 6-wire cable between indicator and junction box in platform | | 282 m/mm ² |

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.