



Member State
Denmark

OIML Certificate N°
R76/2006-DK3-17.16

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**
Address: Venlighedsvej 4
2970 Hørsholm
Denmark

Person responsible: J. Hovgård Jensen

Applicant

Name: **Tscale Electronics Mfg (Kunshan) Co., Ltd.**
Address: No.99 Shunchang Road,
Zhoushi, Kunshan, Jiangsu,
China

Manufacturer

of the certified pattern: **Tscale Electronics Mfg (Kunshan) Co., Ltd.**

Identification

of the certified pattern: **Non-automatic weighing instrument**
Type: IHB600 / IHB6000 / NHB600 / NHB6000 / EHB600 /
EHB6000
characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 2006
for accuracy class II

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.



OIML Certificate N° R76/2006-DK3-17.16

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1913635, dated 03-12-2013 that includes 91 pages

The issuing authority: **DELTA, OIML Issuing Authority DK3**
13 December 2017

J. Hovgård Jensen
Certification Manager

Characteristics

Accuracy class		II
Weighing range		Single-interval
Maximum capacity	Max	600 g or 6000 g
Minimum capacity	Min	20 d
Verification scale interval	e	e = 0.1 g or e = 1 g
Scale interval	d	d = 0.1e
Number of intervals	Max / e	6000
Subtractive tare	T	≤ -Max
Interface		Protective, according to paragraph 5.3.6
Supply voltage		12 VDC / 230 VAC, 50 Hz using external AC/DC
Temperature range for the indicator		+15 °C / +30 °C
Load cell		HBM SPL 600g, HBM PW6K 10 kg
Software identification - legal relevant software: - application software: - optional remote display		1.10 or 1.11 A1.yy 2.10

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.