

Member State of OIML United Kingdom of Great Britain and Northern Ireland OIML Certificate No R76/1992-GB1-04.04

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: National Weights and Measures Laboratory

Address: Stanton Avenue

Teddington Middlesex TW11 0JZ

United Kingdom

Person responsible: Richard Sanders – Assistant Director, Type Approval.

Applicant

Name: Chronos Richardson GmbH

Address: Reutherstrasse 3

D 53773 Hennef

Germany

Identification of the certified pattern:

Chronos Richardson GmbH SpeedAC NXT indicating device Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML: R76
Edition: 1992 (E)
Accuracy class: III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports: No 00457 having 46 pages and associated pattern evaluation checklist F20040 which includes 12 pages.

The issuing authority

The CIML member

Richard Sanders

Jeff Llewellyn

Date 26 February 2004

Date 26 February 2004

Characteristics: This indicating device is designated the SpeedAC NXT indicator. It has the following devices:

- Semi-automatic zero setting device (≤4% of Max)
- Zero-tracking device (≤0.5d/s within 4% Max)
- Subtractive tare device
- Gross and Net Indicator
- Semi automatic tare device
- Pre set tare device
- Display test device
- Time and date function

Comprising of: The indicator housing is fabricated from stainless steel plate. The front panel has a backlit LCD display and a twenty-seven key keyboard, five of the twenty seven keys are programmable software keys.

Technical characteristics of $P_i = 0.5$ (highest precision)

Power supply	115 VAC or 230 VAC
Maximum number of scale intervals	6000
Loadcell excitation voltage	± 5 VDC (10 VDC)
Minimum loadcell impedance	21.875 Ω
Maximum loadcell impedance	2000 Ω
Minimum input voltage per verification scale interval	1 micro volt
Measuring range minimum voltage	-10 mV
Measuring range maximum voltage	70 mV
Fraction of maximum permissible error	$P_{ind} = 0.50$
Operating temperature range	-10°C to +40°C

Loadcell cable	6 cores around PVC filler in centre, tinned copper	
	braid, flexible PVC overall jacket.	
	Maximum length = 100 m for 4-wire operation	

Technical characteristics of $P_i = 0.75$ (lowest precision)

Power supply	115 VAC or 230 VAC	
Maximum number of scale intervals	10000	
Load cell excitation voltage	± 5 VDC (10 VDC)	
Minimum load cell impedance	21.875 Ω	
Maximum load cell impedance	2000 Ω	
Minimum input voltage per verification scale interval	1 micro volt	
Measuring range minimum voltage	-10 mV	
Measuring range maximum voltage	70 mV	
Fraction of maximum permissible error	$P_{ind} = 0.75$	
Operating temperature range	-10°C to +40°C	
Load cell cable	6 cores around PVC filler in centre, tinned copper braid, flexible PVC overall jacket. Maximum length = 100 m for 4-wire operation	

Maximum cable length for 6-wire operation					
Load Cell		Unit of length			
Impedance ¹	0.2 mm ²	0.5 mm^2	$1.0 \mathrm{mm}^2$		
22 Ω	14	33	71	Meters	
44 Ω	28	66	142	Meters	
87 Ω	56	133	283	Meters	
350 Ω	224	535	1134	Meters	

calculated by dividing the single load cell impedance by the number of load cells

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.