



# **OIML Member State** United Kingdom of Great Britain and Northern Ireland

# OIML Certificate No. R76/2006-A-GB1-18.11

OIML CERTIFICATE ISSUED UNDER SCHEME A				
OIML Issuing Authority	NMO Stanton Avenue Teddington TW11 0JZ United Kingdom			
Person responsible:	Mannie Panesar – Head of Technical Services			
Applicant	CAS Corporation #262, Geurugogae-ro Gwangjeok-myeon Yangju-si Gyeonggi-do Republic of Korea			
Manufacturer	The applicant			
Identification of the certified type	<b>RW-5000 Series: RW-5002PL Model</b> (the detailed characteristics are defined in the Descriptive	e Annex)		
sample(s) identified in th	tests the conformity of the above identified type (represent ne OIML type evaluation report) with the requirements of th International Organization of Legal Metrology (OIML):			
OIML R 76-1, Edition:	2006			
For accuracy class: III a				
Issue date: 24 October :	2018			
The OIML Issuing Auth	nority			
Grégory Glas Lead Technical Manag For and on behalf of the He	Joi	UKAS PRODUCT CERTIFICATION 0135		

NMO I Stanton Avenue I Teddington I TW11 OJZ I United Kingdom

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. P02431-4 dated 24 October 2018 that includes 16 pages

The technical documentation relating to the identified type is contained in documentation file:

No. P02431-4-D dated 24 October 2018

# **OIML Certificate History**

Revision No.	Date	Description of the modification
Revision 0	24 October 2018	Certificate first issued.
-	-	-

No revisions have been issued.

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

# DESCRIPTIVE ANNEX

## Characteristics of the instrument:

This instrument, designated the RW-5002PL, utilises the RW-5002PL digital indicating device (part of the RW-5000 Series) connected to two independent weighing platforms (weigh pads) to form a single or dual-interval, Class III or IIII, weighing instrument.

The instrument is portable, self-indicating and battery-powered, and shall not be used for direct sales to the public.

#### Main features:

- Plastic enclosure
- LCD display
- Alphanumerical keypad and function keys
- Integrated printer
- Dual channel (two load cell connections)
- 7-pin load cell connectors

### Devices:

- Initial zero setting device on power up ( $\leq 20\%$  Max)
- Semi-automatic zero setting ( $\leq 4\%$  Max)
- Zero tracking (optional) ( $\leq 4\%$  Max)
- Zero-indicator
- Indication of stable equilibrium
- Gravity compensation
- Printing
- Totalisation of weights
- Load identification register

Note: The summation of axle weights is not permitted under this certificate.

#### Interfaces:

- Load cell connection
- RS232/485
- USB (optional)

## Load cell:

Any compatible load cell(s) may be used providing the following conditions are met:

- There is a respective OIML Certificate of Conformity (R60) issued for the load cell.
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules, and any particular installation requirements. A load cell marked NH is allowed only if humidity testing to R76 has been conducted on this load cell.
- The compatibility of the load cells and indicator is established by the manufacturer by means of the compatibility of modules calculation at the time of verification.
- The load cell transmission conforms to a standard type.

# Technical data (indicator):

Power supply	12 VDC, 1.25 A rechargeable battery	
Maximum number of scale intervals	10,000 (Class III)	
	1,000 (Class IIII)	
Load cell excitation voltage	5 VDC	
Minimum load cell impedance	43.75 Ω	
Maximum load cell impedance	1000 Ω	
Minimum input voltage per verification scale interval	0.5 μV	
Measuring range minimum voltage	0 mV	
Measuring range maximum voltage	16 mV	
Fraction of maximum permissible error	P <sub>i</sub> = 0.5	
Operating temperature range	- 10 °C to + 40 °C	
Load cell cable (from indicator to load cell junction	2 m (4-wire configuration)	
box) - Maximum length	22 m/mm <sup>2</sup> (6-wire configuration)	

## Software:

The software is held in firmware on the circuit board, and has the identification number "V1.xx", with xx reflecting non-legally relevant changes. The software version number is displayed at power-up.

Download of software is only possible by accessing the main board inside the sealed enclosure.

Access to the legally relevant parameters is prevented by a jumper on the main board.

## Sealing:

Access to the electronics, access to the switch described in Software section and the load cell connection are sealed using a tamper-evident method.

#### Alternatives:

Having the instruments manufactured by the following companies:

CAS (Zhejiang) Electronics Co., Ltd 99# Changjiang Road Jiashan County Zhejiang Province China