
OIML Member State Japan	OIML Certificate No. R76/2006-A-JP1-19.02
OIML CERTIFICATE ISSUED UNDER SCHEME A	
OIML Issuing Authority Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST) Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan Person responsible: Dr. Ryoji Chubachi, President of AIST	
Applicant Name: A&D Company, Limited. Address: 3-23-14 Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN	
Manufacturer Name: Litra CO., LTD Address: 7-5, Harajyuku, Hidaka-shi, Saitama 350-1205, JAPAN Name: A&D SCALES CO., LTD. Address: 191, Inseok-ro, Deoksan-myeon, Jincheon-gun, Chungcheongbuk-do, 27856, KOREA	
Identification of the certified type (the detailed characteristics will be defined in the additional pages) Models: AD-6106 series, AD-6107 series	
Designation of the module (if applicable) Non-automatic weighing instruments	
<p>This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):</p> <p>OIML R 76-1, Edition: 2006</p> <p>For accuracy class: (III)</p>	

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 31-001, dated 4 July 2019, that includes 17 pages

The technical documentation relating to the identified type is contained in documentation file:

No. 31-001-D, dated 4 July 2019

OIML Certificate History

Revision No.	Date	Description of the modification
Revision 0	29 July 2019	OIML Certificate first issued
-	-	-
-	-	-
-	-	-
-	-	-
-	-	-


This revision replaces previous versions of the certificate.

Identification, signature and stamp

The Issuing Authority

NMIJ/AIST

The OIML Member



Dr. R. Chubach
President of AIST
29 July 2019

Dr. Y. Miki
29 July 2019

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

DESCRIPTIVE ANNEX

Characteristics of the instrument:

The AD-6106R and AD-6107R series is a class III, self-indicating, non-automatic weighing instrument. The instruments are not designed for direct sales to the public.

Technical data:

Type	AD-6106R	AD-6107R
Class	III	
Max	150/200 kg	
e	0.1/0.2 kg	
n	1500/1000	
Min	2 kg	
Maximum tare-balancing	90 kg	
Preset tare	90 kg	
Temperature range	+ 5 to + 35 °C	
Power supply	AC adapter and/or Dry battery and/or Rechargeable battery	AC adapter and/or Dry battery

Device:

- Initial zero-setting device ($\leq 10\%$ of Max)
- Semi-automatic zero-setting device ($\leq 4\%$ of Max)
- Zero-tracking device ($\leq 4\%$ of Max)
- Semi-automatic subtractive tare weighing device
- Preset tare device
- Zero indicator
- Indication of stable equilibrium device

Interfaces:

- Serial data interface RS232C (to connect Printer)
- Bluetooth (to connect Personal computer)

Software:

The legally relevant software is designated version P-1.xx, with x reflecting non-legally relevant changes.

Sealing:

Access to the load cell, electronics and calibration switch is prevented by sealing the enclosure using a tamper-evident method.