



OIML Member State
Japan

OIML Certificate No.
R76/2006-A-JP1-19.03
Revision 2

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3, 1-1-1 Umezono Tsukuba Ibaraki 305-8563, Japan

Person responsible: ISHIMURA Kazuhiko, President of AIST

Applicant

Name: Paramount Bed Co., Ltd.
Address: 14-5, 2-chome, Higashisuna, Koto-ku, Tokyo, 136-8670, Japan

Manufacturer

Name: Paramount Bed Co., Ltd.
Address: 14-5, 2-chome, Higashisuna, Koto-ku, Tokyo, 136-8670, Japan

Identification of the certified type

(the detailed characteristics will be defined in the additional pages)

Non-automatic weighing instruments
Models: SB series

Designation of the module (if applicable)

N/A

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1, Edition: 2006

For accuracy class: III

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 2021-010, dated 25 March 2022, that includes 5 pages

The technical documentation relating to the identified type is contained in documentation file:

No. 2021-010-D, dated 25 March 2022

OIML Certificate History

Revision No.	Date	Description of the modification
Revision 0	27 September 2019	OIML Certificate first issued
Revision 1	12 June 2020	Indicator and Sealing added
Revision 2	29 March 2022	The company changed its address The technical documentation is revised

This revision replaces previous versions of the certificate.

Identification, signature and stamp

The Issuing Authority
NMIJ/AIST

The OIML Member



ISHIMURA Kazuhiko
President of AIST
29 March 2022

TAKATSUJI Toshiyuki
29 March 2022

The accreditation body:

NMIJ/AIST has achieved accreditation under the ASNITE-Product (OIML-CS) scheme of IAJapan, which applies ISO/IEC 17065:2012 and regulations relevant to OIML-CS as the accreditation criteria. The accreditation identification for this accreditation is ASNITE 0001 Product and the details of the accreditation information could be referred from the IAJapan website (<https://www.nite.go.jp/en/iajapan/asnite/lab/index.html>).

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

DESCRIPTIVE ANNEX

Characteristics of the instrument:

The SB...series is a class III, self-indicating, non-automatic weighing instrument. The instruments are not designed for direct sales to the public.

Technical data:

Type	SB...series
Class	III
Max	255 kg
e	0.1 kg
Min	2 kg
Temperature range	+ 5 °C to + 40 °C

Device:

- Semi-automatic zero-setting device ($\leq 4\%$ of Max)
- Semi-automatic subtractive tare weighing device
- Zero indicator
- Indication of stable equilibrium device

Interfaces:

Software:

The legally relevant software is designated version V.xx.01.xx (Scale BOX) and V.xx.03.xx (Terminal), with x reflecting non-legally relevant changes.

Sealing:

Access to the load cell, electronics and calibration switch is prevented by a tamper-evident seal on the base of the instrument.