

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Teraoka Weigh-System PTE Ltd.
Address: 4 Leng Kee Road, #05-03/04/05 & 11
SIS Building, 159088
Republic of Singapore

Manufacturer of the certified type

Name: Teraoka Weigh-System PTE Ltd.
Address: 4 Leng Kee Road, #05-03/04/05 & 11
SIS Building, 159088
Republic of Singapore

Identification of certified type

Non-automatic weighing instrument
Type: SM-50xx, SM-54xx, SM-55xx
 $6 \text{ kg} \leq \text{Max} \leq 300 \text{ kg}$
 $e \geq 1 \text{ g}$
 $n \leq 3000$ divisions (per partial weighing range)
 $n \leq 7500$ divisions (for single-interval instruments)
maximum of 2 partial weighing ranges
 $T \leq - \text{Max}_1$
Temperature range -10 °C / 40 °C



This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7584 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 2006 (E)
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:
N° R76/2006-NL1-09.19A that includes 52 pages;
N° R76/2006-NL1-09.19B that includes 35 pages.

The Issuing Authority NL1
NMI Certin, 28 May 2009


C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.