

OIML Member stateThe Netherlands

OIML Certificate N° R76/2006-NL1-10.02

Project number 9200687 Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMi Certin B.V.

Address: Hugo de Grootplein 1

3314 EG Dordrecht The Netherlands

Person responsible: Ing. C. Oosterman

Applicant

Name: + + + Mettler-Toledo (Changzhou) Measurement Technology Ltd.

Address: 111, West HaiHu Rd, XinBei District,

ChangZhou, 213125 Jiangsu

Peoples Republic of China

Manufacturer of the certified type

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.

Address: 111, West HaiHu Rd, XinBei District,

ChangZhou, 213125 Jiangsu

Peoples Republic of China

Identification of certified type

Indicator, as a part of a non-automatic weighing instrument.

Type: IND131 / IND331

The maximum number of verification scale intervals will be:

 $n \le 6000$ for class 1 instruments or; $n \le 1000$ for class 1 instruments.

Applied error fraction p.: 0.5 Minimum input impedance: 87 Ω Maximum input impedance: 1200 Ω

Signal voltage per verification scale interval ≥ 0.6 µV

Temperature range -10°C / +40°C

NMi Certin BV

Hugo de Grootplein 1, 3314 EG Dordrecht PO Box 394, 3300 AJ Dordrecht, NL T+31 78 6332332 F+31 78 6332309 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.
The notification of NMi Certin as Issuing Authority can be verified at www.oiml.org.





OIML Member state

The Netherlands

OIML Certificate N° R76/2006-NL1-10.02

Project number 9200687 Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC7600 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 2006 (E)

for accuracy class III or III

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/2006-NL1-09.26A that includes 38 pages; N° R76/2006-NL1-09.26B that includes 30 pages;

N° R76/2006-NL1-10.02 that includes 19 pages.

The Issuing Authority NL1 NMi Certin, 8 February 2010

C. Oosterman

Head Certification Board

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although

either may be reproduced in full.