

OIML Member state The Netherlands

OIML Certificate N° R76/2006-NL1-10.05

Project number 9200745 Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:

NMi Certin B.V.

Address:

Hugo de Grootplein 1 3314 EG Dordrecht

The Netherlands

Person responsible:

Ing. C. Oosterman

Applicant

Name:

Grupo Epelsa, S.L.

Address:

Ctra. Sta. Cruz de Calafell 35

E-08830 Sant Boi de Llobregat (Barcelona)

Spain

Manufacturer of the certified type

Name:

Grupo Epelsa, S.L.

Address:

Ctra. Sta. Cruz de Calafell 35

E-08830 Sant Boi de Llobregat (Barcelona)

Spain

Identification of certified type

Electronic, self-indicating device, with single [- or multi]-interval or multi range

indication.

Type: Energy

The maximum number of verification scale intervals will be:

n ≤ 7500 for class (III) instruments or

 $n_i \leq 3000$ per partial weighing range for class $\stackrel{\hbox{$1$}{\hbox{$1$}}}{\hbox{$1$}}$ multi-interval and multi-range

instruments (with a maximum of two partial weighing ranges) or

 $n \le 1000$ for class (III) instruments.

Applied error fraction pi: 0.5.

NMi Certin BV
Hugo de Grootplein 1, 3314 EG Dordrecht
PO Box 394, 3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted. The notification of NMi Certin as Issuing Authority can be verified at www.oiml.org.







OIML Member state

The Netherlands

OIML Certificate N° R76/2006-NL1-10.05

Project number 9200745 Page 2 of 2

Minimum input impedance: 35 Ω Maximum input impedance: 1050 Ω

Signal voltage per verification scale interval ≥ 0.66 μV

Temperature range -10 °C / +40 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC7677 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 2006 (E)

for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/2006-NL1-10.05 that includes 49 pages.

The Issuing Authority NL1 NMi Certin, 8 March 2010

. Oosterman

Head Certification Board

* *

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.