



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-12.11  
Project number 11200063  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Mettler-Toledo (ChangZhou) Measurement Technology Ltd. No.111, West TaiHu Road, Xin Bei Discript, ChangZhou, 213125, Jiangsu, Peoples Republic of China
Manufacturer	Mettler-Toledo (ChangZhou) Measurement Technology Ltd. No.111, West TaiHu Road, Xin Bei Discript, ChangZhou, 213125, Jiangsu, Peoples Republic of China
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : bPro.../ bC-.../ RL20-...
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
25 May 2012

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-12.11  
Project number 11200063  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- N° R76/1992-NL1-06.18A that includes 44 pages;
- N° R76/1992-NL1-06.18B that includes 28 pages;
- N° R76/1992-NL1-07.11A that includes 16 pages;
- N° R76/1992-NL1-07.11B that includes 26 pages;
- N° R76/1992-NL1-09.30 that includes 12 pages;
- N° 811744 that includes 12 pages;
- N° NMI-12200063-03 that includes 26 pages.

**Characteristics of the non-automatic weighing instrument:**

- $6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
- $n \leq 6000$  divisions in case of single interval and  $e \geq 2 \text{ g}$
- $n \leq 3000$  division (per weighing range) in case of multi-interval and  $e \geq 1 \text{ g}$

Maximum of 2 partial weighing ranges

Temperature range  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

In case of the instrument using scale bases covered by a Test Certificate using the IDNet interface :  $6 \text{ kg} \leq \text{Max} \leq 1500 \text{ kg}$