

OIML Member State
The Netherlands

Number R76/2006-NL1-12.18
Project number SO12200135
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Hiweigh Technologies Ltd. No.2278, Zhaotai Road Puijang Town, Minhang District, Shanghai 201112 China
Manufacturer	Hiweigh Technologies Ltd. No.2278, Zhaotai Road Puijang Town, Minhang District, Shanghai 201112 China
Identification of the certified type	An indicator as a part of a non-automatic weighing instrument Type : X3AM X3M X5M X1M
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
21 May 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-12.18
Project number SO12200135
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number NMI-11200266-01 dated 6 October 2011 that includes 34 pages;
Number NMI-11200266-02 dated 6 October 2011 that includes 31 pages.

Characteristics of the indicator:

Type: X3AM, X3M, X5M, X1M

The maximum number of verification scale intervals will be:

$n \leq 3000$ for class III instruments or;

$n \leq 1000$ for class III instruments.

Applied error fraction p_i : 0,5

Minimum input impedance: 87Ω

Maximum input impedance: 1050Ω

Temperature range: $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

Signal voltage per verification scale interval $\geq 1 \mu\text{V}$