

OIML Member State
The Netherlands

Number R76/2006-NL1-12.27
Project number 12200160
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Tanita Corporation 14-2,1 chome, Maeno-cho Itabashi-ku Tokyo, 174-8630
Manufacturer	Tanita Corporation 14-2,1 chome, Maeno-cho Itabashi-ku Tokyo, 174-8630
Identification of the certified type	A Non-automatic weighing instrument Type : WB-150
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
30 July 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-12.27
Project number 12200160
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number NMI-12200160-01 dated 30 July 2012 that includes 54 pages.

Characteristics of the non-automatic weighing instrument:

- Maximum capacity : $\text{Max} \leq 270 \text{ kg}$
- Verification scale interval : $e \geq 100 \text{ g}$
- Maximum number of scale intervals : $n \leq 2700 \text{ divisions}$
- Temperature range : $+5 \text{ }^\circ\text{C} / +35 \text{ }^\circ\text{C}$
- Tare : $T \leq -20 \text{ kg}$
- Software version number : 9351