

OIML Member State
The Netherlands

Number R76/2006-NL1-12.39
Project number 11200555
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Shanghai Teraoka Electronic Co., Ltd. Tinglin Industry Developmental Zone Jinshan District, Shanghai 201505 Peoples Republic of China
Manufacturer	Shanghai Teraoka Electronic Co., Ltd. Tinglin Industry Developmental Zone Jinshan District, Shanghai 201505 Peoples Republic of China
Identification of the certified type	A Non-automatic weighing instrument Type : DS-425
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
21 November 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

OIML Member State
The Netherlands

Number R76/2006-NL1-12.39
Project number 11200555
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL-02.49 dated 19 December 2002 that includes 52 pages;
- No. R76/1992-NL-03.35 dated 17 October 2003 that includes 53 pages;
- No. NMI-11200555-01 dated 21 November 2012 that includes 49 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$0,6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 0,1 \text{ g}$
Maximum number of scale intervals	$n \leq 6000$ divisions
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	$T \leq -50\%$ of Max
Weighing range(s)	Single interval
Power supply voltage	220 –240 V AC 50/60 Hz 18 V DC
Software identification	Version number: 2.xx for the DS-425 and DS-425P 1.xx for the DS-425P(V1) Where xx is a number between 00 and 99 which presents the non-legally relevant software.