



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.02  
Project number 12200765  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Mettler-Toledo (Changzhou) Measurement Technology Ltd. No. 111 West Taihu Road, ChangZhou XinBei District 213125 Jiangsu, Peoples Republic of China
Manufacturer	Mettler-Toledo (Changzhou) Measurement Technology Ltd. No. 111 West Taihu Road, ChangZhou XinBei District 213125 Jiangsu, Peoples Republic of China
Identification of the certified type	An <b>Indicator</b> Type : IND226x
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 2006 for accuracy class **III** and **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
27 February 2013

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).



**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.02  
Project number 12200765  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. 509535 dated 13 December 2005 that includes 44 pages;
- No. 509535A dated 13 December 2005 that includes 16 pages;
- No. 605617 dated 8 September 2006 that includes 35 pages;
- No. 700235A dated 9 March 2007 that includes 16 pages;
- No. 700235B dated 9 March 2007 that includes 14 pages;
- No. NMI-12200765-01 dated 21 February 2012 that includes 24 pages.

**Characteristics of the indicator:**

Accuracy class	III and IIII
Maximum number of verification scale intervals	6000
Load cell excitation voltage	5 V DC
Minimum input voltage per verification scale interval	0,8 $\mu$ V
Minimum load cell resistance	87 $\Omega$
Maximum load cell resistance	1100 $\Omega$
Temperature range	-10 $^{\circ}$ C / +40 $^{\circ}$ C
Fraction of the maximum permissible error	0,5
Load cell connection	4-wire 6-wire (remote sensing)
Maximum value of the cable length per cross wire section (6-wire system)	No special cable length has to be provided for the connection between the indicator and the junction box or load cells.
Weighing range(s)	Single interval
Power supply voltage	100 – 240 V AC 50/60 Hz
Maximum number of load platforms	1
Software identification	Version number: 198005 L1.xx, where the xx part represents the non-essential software modifications