



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-13.07 revision 1
Project number 13200101
Page 1 of 3

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Manufacturer	Mettler-Toledo (Changzhou) Measurement Technology Ltd. 111 West Taihu Road XinBei District, Changzhou Jiangsu 213125 Peoples Republic of China
Identification of the certified type	A Non-automatic weighing instrument Type : IND211 / XIG / IND220 / IND221 / IND226 / IND226x / IND231 / IND236 BBA211 / BBA220 / BBA221 / BBA226 / BBA226x / BBA23x -yyyyyyyy x defines the explosion proof version (optional) x-yyyyyyyy, defined by the manufacturer.
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class $\textcircled{\text{III}}$ or $\textcircled{\text{III}}$

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
7 May 2013

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-13.07 revision 1
Project number 13200101
Page 2 of 3

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- N° 509535 that includes 44 pages;
- N° 509535A that includes 16 pages;
- N° 605617 that includes 35 pages;
- N° 700235A, that includes 16 pages;
- N° 700235B, that includes 14 pages;
- N° R76/2006-NL1-08.35a that includes 29 pages;
- N° R76/2006-NL1-08.35b that includes 25 pages;
- N° R60/1991-NL-98.02A, that includes 36 pages;
- N° R60/1991-NL-98.02B, that includes 37 pages;
- N° R60/1991-NL-98.16A, that includes 38 pages;
- N° R60/1991-NL-98.16B, that includes 35 pages;
- N° R60/2000-NL-01.03, that includes 40 pages;
- N° R60/2000-NL-01.06A, that includes 40 pages;
- N° R60/2000-NL-01.06B, that includes 37 pages;
- N° R60/2000-NL-02.10A, that includes 40 pages;
- N° R60/2000-NL-02.10B, that includes 37 pages;
- N° R60/2000-NL-02.11A, that includes 40 pages;
- N° R60/2000-NL-02.11B, that includes 37 pages;
- N° R60/2000-NL-02.12, that includes 40 pages;
- N° R60/2000-NL-02.13, that includes 40 pages;
- N° R60/2000-NL-02.14A, that includes 40 pages;
- N° R60/2000-NL-02.14B, that includes 37 pages;
- N° R60/2000-NL-02.17A, that includes 43 pages;
- N° R60/2000-NL-02.17B, that includes 37 pages;
- N° R60/2000-NL-02.29, that includes 40 pages;
- N° R60/2000-NL1-03.10A, that includes 37 pages;
- N° R60/2000-NL1-03.10B, that includes 38 pages;
- N° R60/2000-NL1-03.14A, that includes 37 pages;
- N° R60/2000-NL1-03.14B, that includes 40 pages;
- N° R60/2000-NL1-03.24, that includes 37 pages;
- N° R60/2000-NL1-04.05, that includes 40 pages;
- N° R60/2000-NL1-04.12A, that includes 40 pages;
- N° R60/2000-NL1-04.12B, that includes 37 pages;
- N° R60/2000-NL1-04.12C, that includes 37 pages;
- N° R60/2000-NL1-04.15A, that includes 40 pages;
- N° R60/2000-NL1-04.15B, that includes 37 pages;
- N° R60/2000-NL1-04.19A, that includes 43 pages;
- N° R60/2000-NL1-04.19B, that includes 37 pages;
- N° R60/2000-NL1-04.41, that includes 38 pages;
- N° R60/2000-NL1-05.02A, that includes 38 pages;

OIML Member State
The Netherlands

Number R76/2006-NL1-13.07 revision 1
Project number 13200101
Page 3 of 3

- N° R60/2000-NL1-05.02B, that includes 37 pages;
- N° R60/2000-NL1-05.09, that includes 40 pages;
- N° R60/2000-NL1-05.18A, that includes 38 pages;
- N° R60/2000-NL1-05.18B, that includes 37 pages;
- N° R60/2000-NL1-05.18C, that includes 40 pages.
- N° NMI-13200101-01 that includes 49 pages;
- N° NMI-13200101-02 that includes 13 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	1 kg ≤ Max ≤ 10.000 kg
Verification scale interval	e ≥ 1 g
Maximum number of scale intervals	n ≤ 3000 divisions (per partial weighing range)
Temperature range	-10 °C / +40 °C

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
Initial	6 May 2013	-
1	7 May 2013	Change of number in type