

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.08  
Project number 12200757  
Page 1 of 2

|                                      |   |
|--------------------------------------|---|
| Issuing authority                    | NMi Certin B.V.<br>Person responsible: C. Oosterman   |
| Applicant                            | Shanghai Teraoka Electronic Co., Ltd.<br>Tinglin industry developmental zone<br>Jin Shan county, Shanghai 201505<br>P.R. of China |
| Manufacturer                         | Shanghai Teraoka Electronic Co., Ltd.<br>Tinglin industry developmental zone<br>Jin Shan county, Shanghai 201505<br>P.R. of China |
| Identification of the certified type | <b>A Non-automatic weighing instrument</b><br>Type : DS-620, DS-620SS   |
| Characteristics                      | See next page   |

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 2006 for accuracy class **III** or **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
18 April 2013

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl)).

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. 807678 dated 4 December 2008 that includes 44 pages;
- No. NMI-12200757-01 dated 15 April 2013 that includes 28 pages;
- No. NMI-12200757-02 dated 15 April 2013 that includes 13 pages;
- R60/1991-NL-97.14, that includes 37 pages (load cell type P);
- R60/1991-NL-99.09, that includes 36 pages (load cell type P);
- R60/1991-NL-97.23A, that includes 37 pages (load cell type N);
- R60/1991-NL-97.23B, that includes 37 pages (load cell type N);
- R60/1991-NL-97.18A, that includes 37 pages (load cell type K);
- R60/1991-NL-97.18B, that includes 37 pages (load cell type K);
- R60/1991-NL-95.07, that includes 38 pages (load cell type M).

**Characteristics of the non-automatic weighing instrument:**

|                                   |   |
|-----------------------------------|---|
| Accuracy class                    | III or IIII   |
| Maximum capacity                  | $6 \text{ kg} \leq \text{Max} \leq 600 \text{ kg}$  |
| Verification scale interval       | $e \geq 1 \text{ g}$  |
| Maximum number of scale intervals | $n \leq 7500$ divisions<br>(per partial weighing range)   |
| Maximum partial weighing ranges   | 2   |
| Temperature range                 | $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$   |
| Tare                              | $T \leq -50\%$ for single interval instruments<br>$T \leq -\text{Max}_1$ for multi interval instruments |
| Weighing range(s)                 | Single interval<br>Multi-interval   |
| Power supply voltage              | 220 – 240 V AC 50/60 Hz<br>Battery 6 V  |
| Software identification           | Version number: 1.xx or 2.xx<br>(xx reserved for non-essential changes)                                 |