



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-13.13 Revision 2
Project number 12200731
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer Xiamen Pinnacle Electrical Co., Ltd.
4F, Guangxia Building, Torch High-Tech Zone
361006 Xiamen, Fujian
China

A Non-automatic weighing instrument

Type : OS2 series

Brand : ACLAS or Arm Pos

Manufacturers name : Xiamen Pinnacle Electrical Co., Ltd.

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
19 September 2017


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-13.13 Revision 2
Project number 12200731
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-10200906-01 dated 20 April 2011 that includes 36 pages;
- No. NMI-12200731-01 Revision 1 dated 1 September 2017 that includes 35 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Maximum number of scale intervals	$n \leq 3000$ divisions (per partial weighing range)
Maximum partial weighing ranges	2
Temperature range	0 °C / +40 °C
Tare	$T \leq -(\text{Max}_1 - e_1)$ for multi-interval instruments
Weighing range(s)	Multi-interval
Power supply voltage	100 – 240 V AC 50/60 Hz 5 V DC
Application	Intended to be used for direct sales to the public

Revision History

This revision replaces the previous versions.

Revision	Date	Change(s)
Initial	23 May 2013	-
1	27 May 2013	Typing error. Type name missing by "Type"
2	19 September 2017	Typing errors. References to project number and test reports corrected.