

OIML Member State
The Netherlands

Number R76/2006-NL1-13.17
Project number 12200013
Page 1 of 2

Issuing authority	NMI Certin B.V. Person responsible: C. Oosterman	
Applicant	CAS Corporation #19, Ganap-ri, Gwangjuk-Myoun Yangju-Si, Geonggi-do South Korea	
Manufacturer	CAS Corporation #19, Ganap-ri, Gwangjuk-Myoun Yangju-Si, Geonggi-do South Korea	
Identification of the certified type	A Indicator Type	: CI-200D, CI-201D, CI-200SD, CI-200SCD, CI-201SD
Characteristics	See next page	

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **(III)** and **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
31 May 2013

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-13.17
Project number 12200013
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-12200013-01 dated 24 May 2013 that includes 46 pages;
- No. NMI-12200013-02 dated 24 May 2013 that includes 16 pages.

Characteristics of the indicator:

Accuracy class	III and IIII
Temperature range	-10 °C / +40 °C
Fraction of the maximum permissible error	0
Weighing range(s)	Single interval
Power supply voltage	110-240 V AC 50/60 Hz
Maximum number of digital load cells	8
Software identification	Version number: U 2.04
Power supply to digital load cell	9 V DC