



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-13.28
Project number 11200884
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Biospace Co., Ltd 15, Heugam-gil, Ipjang-myeon, Seobuk-gu, Cheonan-si, Chungcheongnam-do 331-824, South-Korea
Manufacturer	Biospace Co., Ltd 15, Heugam-gil, Ipjang-myeon, Seobuk-gu, Cheonan-si, Chungcheongnam-do 331-824, South-Korea
Identification of the certified type	A Non-automatic weighing instrument Type : InBody 770
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
2 July 2013

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).



OIML Member State
The Netherlands

Number R76/2006-NL1-13.28
Project number 11200884
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-10200704-01 dated 25 October 2011 that includes 53 pages;
- No. NMI-11200884-01 dated 28 June 2013 that includes 46 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max \leq 270 kg
Verification scale interval	$e \geq$ 100 g
Maximum number of scale intervals	$n \leq$ 2700 divisions
Temperature range	+5 °C / +35 °C
Weighing range(s)	Single interval
Power supply voltage	Adapter 100 – 240 V AC 50/60 Hz to 12 V DC
Application	Intended to be used for weighing patients
Software identification	Version number: "IB770-1.0(Lock)"