OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-13.32
Project number 11200389
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Precisa Gravimetrics AG Moosmattstrasse 32 CH-8953 Dietikon Switzerland
Identification of the certified type	A **Non-automatic weighing instrument** Type : Serie LX/LS
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class Ⓘ and Ⓘ Ⓘ

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority	**NMi Certin B.V., OIML Issuing Authority NL1** 10 September 2013 C. Oosterman Head Certification Board

OIML MUTUAL ACCEPTANCE ARRANGEMENT

RvA INSPECTION RvA I 122

OIML Certificate of Conformity

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMi-11200389-01 dated 12 August 2013 that includes 37 pages;
- No. NMi-11200389-02 dated 12 August 2013 that includes 29 pages;
- No. NMi-11200389-03 dated 12 August 2013 that includes 17 pages;
- No. NMi-11200389-04 dated 12 August 2013 that includes 17 pages.

Characteristics of the non-automatic weighing instrument:

	I	II
Accuracy class	I	II
Maximum capacity	$120\ g \leq Max \leq 1220\ g$	$160\ g \leq Max \leq 10200\ g$
Verification scale interval	$e \geq 1\ mg$	$e \geq 10\ mg$
Maximum number of scale intervals	$n \leq 320000$ divisions (per partial weighing range)	
Maximum partial weighing ranges	1	
Temperature range	$+15\ °C\ /\ +25\ °C$	$+10\ °C\ /\ +30\ °C$
Tare	$T \leq -Max$	
Weighing range(s)	Single interval	
Power supply voltage	100 – 240 V AC 50/60 Hz (with external adapter) 12 V DC (supplied by external adapter)	
Software identification	Version number: xx.xx ExxD01 where xx can be 00-99 or A-Z, represents the non legally relevant parts. E..D01. represent the legal-relevant part of the software	