

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.37  
Project number 13200392  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo AG Im Langacher CH 8606 Greifensee Switzerland
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : XPE, XSE and XVE
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76-1**, Edition 2006 for accuracy class **I**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
15 October 2013

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-13.37  
Project number 13200392  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200392-01 dated 10 October 2013 that includes 39 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	I
Maximum capacity	Max $\leq$ 220 g
Verification scale interval	e = 1 mg
Maximum number of scale intervals	n $\leq$ 220 000 divisions
Actual scale interval	e = d, or e = 10 d, or e = 20 d, or e = 50 d, or e = 100 d
Temperature range	+10 °C / +30 °C
Tare	T $\leq$ -Max
Weighing range(s)	Single interval
Power supply voltage	100 V –240 V AC 50/60 Hz
Software identification Weigh Bridge	Version number: 1.xx (where xx represents metrological not relevant changes)