

OIML Member State
The Netherlands

Number R76/2006-NL1-13.39
Project number 13200410
Page 1 of 2

| | |
|--------------------------------------|---|
| Issuing authority | NMi Certin B.V. Person responsible: C. Oosterman |
| Applicant and Manufacturer | Dibal S.A. Astintze Kalea, 24 - Pol. Ind. Neinver 48160 Derio, Vizcaya Spain |
| Identification of the certified type | An Indicator Type : 500, 500-SW and D-900 Series |
| Characteristics | See next page |

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76-1, Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
15 October 2013



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).



OIML Member State
The Netherlands

Number R76/2006-NL1-13.39
Project number 13200410
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No.R76/2006-NL1-10.28A dated 26 August 2010 that includes 47 pages;
- No.R76/2006-NL1-10.28B dated 26 August 2010 that includes 16 pages;
- No.R76/2006-NL1-10.46 dated 21 December 2010 that includes 17 pages;
- No.NMi-10201100-01 dated 27 April 2011 that includes 35 pages;
- No.NMi-11200395-01 dated 22 July 2011 that includes 32 pages;
- No.NMi-11200653-01 dated 30 November 2011 that includes 37 pages;
- No.NMi-12200102-01 dated 23 July 2012 that includes 14 pages;
- No.NMi-12200562-01 dated 11 December 2012 that includes 30 pages;
- No. NMi-13200410-01 dated 9 October 2013 that includes 17 pages;
- No. NMi-13200410-02 dated 9 October 2013 that includes 10 pages.

Characteristics of the indicator:

| | |
|--|--|
| Accuracy class | Ⓜ and ⓂⓂ |
| Maximum number of verification scale intervals | 6000 |
| Load cell excitation voltage | 5 V DC |
| Minimum input voltage per verification scale interval | 1,0 μV |
| Minimum load cell resistance | 300 Ω |
| Maximum load cell resistance | 900 Ω |
| Temperature range | -10 °C / +40 °C |
| Fraction of the maximum permissible error | 0,5 |
| Load cell connection | 4-wire 6-wire (remote sensing) |
| Maximum value of the cable length per cross wire section (6-wire system) | The load cell cable or cables shall be connected directly to the indicator without a junction box. |
| Weighing range(s) | Single interval Multi-interval Multiple range |
| Power supply voltage | 110 – 230 V AC 50/60 Hz or 12-24 V DC via external power supply or internal battery |
| Maximum number of load platforms | 1 |
| Application | Intended to be used for direct sales to the public |
| Software identification | 1.00 for the 500 Series; 2.00 for the 500-SW Series. |