

OIML Member State
The Netherlands

Number R76/2006-NL1-14.06
Project number SO14200688
Page 1 of 2

| | |
|--------------------------------------|--|
| Issuing authority | NMi Certin B.V. Person responsible: C. Oosterman |
| Applicant and Manufacturer | Precisa Gravimetrics AG Moosmattstrasse 32 CH-8953 Dietikon Switzerland |
| Identification of the certified type | A Non-automatic weighing instrument Type : Serie LX/LS/LT |
| Characteristics | See next page |

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **I** and **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
17 February 2014



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-11200389-01 dated 12 August 2013 that includes 37 pages;
- No. NMI-11200389-02 dated 12 August 2013 that includes 29 pages;
- No. NMI-11200389-03 dated 12 August 2013 that includes 17 pages;
- No. NMI-11200389-04 dated 12 August 2013 that includes 17 pages.

Characteristics of the non-automatic weighing instrument:

| Accuracy class | I | | II |
|-----------------------------------|---|---------------------------------|--|
| Maximum capacity | 120 g ≤ Max ≤ 1220 g | | 160 g ≤ Max ≤ 10200 g |
| Verification scale interval | e ≥ 1 mg | | e ≥ 10 mg |
| Maximum number of scale intervals | n ≤ 320000 divisions (per partial weighing range) | | |
| Maximum partial weighing ranges | 1 | | |
| Temperature range | +15 °C / +25 °C | | +10 °C / +30 °C |
| Tare | T ≤ -Max | | |
| Weighing range(s) | Single interval | | |
| Power supply voltage | 100 – 240 V AC 50/60 Hz (with external adapter) 12 V DC (supplied by external adapter) | | |
| Software identification | Type LX/LS | Version number: xx.xx ExxD01 | where xx can be 00-99 or A-Z, represents the non legally relevant parts and E..D00/E..D01 represent the legal- relevant part of the software |
| | Type LT | Version number: xx.xx ExxD00 | |