

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-14.12  
Project number SO14201305  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Ohaus Corporation 7 Campus Drive, Suite 310 Parsippany, NJ 07054 United States of America
Identification of the certified type	A <b>Non-automatic weighing instrument</b> Type : RE series, Aviator 3000 A32P... series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
29 April 2014



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-14.12  
Project number SO14201305  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-03.11 dated 7 April 2003 that includes 56 pages;
- No. R76/1992-NL1-07.13 dated 7 May 2007 that includes 36 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ , or $6 \text{ lb} \leq \text{Max} \leq 60 \text{ lb}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions
Temperature range	$-10^\circ\text{C} / +40^\circ\text{C}$
Power supply voltage	AC/DC adapter; input: 100 – 240 V AC 50/60 Hz, output: 9 V - 12V DC Battery; 9V DC
Application	Intended to be used for direct sales to the public
Software identification	Version number: L 1.7