



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-14.26  
Project number 14200270  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Balanças Marques de José Pimenta Marques, Lda. Parque Industrial de Celeirós (2a Fase) Apartado 2376 4701-905 Braga Portugal
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : BM5
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
3 July 2014



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).



**OIML Member State**  
The Netherlands

Number R76/2006-NL1-14.26  
Project number 14200270  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-14200270-01 dated 27 June 2014 that includes 60 pages;
- No. NMI-14200270-02 dated 27 June 2014 that includes 14 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range and multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -\text{Max}$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	230 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Version number ETBAL: 1.0 Checksum: 0x76b2