



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-14.30
Project number 12200108
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome, Ohta-Ku, Tokyo 146-8580, Japan
Identification of the certified type	A Non-automatic weighing instrument Type : DPS-5600
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
22 August 2014


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-12200108-01 dated 25 June 2014 that includes 49 pages;
- No. NMI-12200108-03 dated 10 October 2012 that includes 18 pages;
- No. NMI-12200108-04 dated 10 October 2012 that includes 18 pages;
- No. NMI-12200108-05 dated 10 October 2012 that includes 18 pages;
- No. NMI-12200108-06 dated 10 October 2012 that includes 18 pages;
- No. NMI-12200108-13 dated 13 May 2013 that includes 15 pages;
- No. R76/1992-NL1-10.11 dated 16 March 2010 that includes 26 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 150 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -\text{Max}$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Application	Intended to be used for the making-up of prepackages
Software identification console	Version number 1.xx Where xx is a number between 00 and 99 that represents minor versions that contain bug fixes and non-legally relevant changes
Software identification A/D-board	Version number: 3.xx Where xx is a number between 22 and 99 that represents minor versions that contain bug fixes and non-legally relevant changes