

OIML Member State
The Netherlands

Number R76/2006-NL1-14.37 Revision 1
Project number SO16202511
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Dräger Medical Systems, Inc. 3135 Quarry Road Telford, PA 18969 United States of America
Identification of the certified type	A Non-automatic weighing instrument Type : Isolette 8000 Isolette 8000 plus C2000 Isolette Infant Incubator
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
7 November 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-14.37 Revision1
Project number SO16202511
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report:

- No. NMI-14200170-01 dated 11 September that includes 53 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max \leq 7 kg
Verification scale interval	e \geq 10 g
Weighing range(s)	Single interval
Maximum number of scale intervals	n \leq 700 divisions
Temperature range	+10 °C / +40 °C
Power supply voltage	230 V AC 50/60 Hz
Application	To be used only for weighing patients
Software identification	Version numbers: 3.xx for C2000 Isolette Infant incubator, 4.xx for Isolette 8000, 5.xx for Isolette 8000 plus. (xx is a number between 00...99)

Revision History

This revision replaces the previous version.

Revision	Date	Changes
0	14 September 2014	Initial issue
1	7 November 2016	Model name "Isolette 8000 plus" added, model name C2000 changed to "C2000 Isolette Infant Incubator", software identification detailed for each model