

OIML Member State
The Netherlands

Number R76/2006-NL1-14.41
Project number 13200647
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer GRUPO EPELSA, S.L.
c/. Punto Net, 3
Polígono Tecnológico TECNOALCALÁ
E-28805 Alcalá de Henares
Madrid
Spain

Identification of the certified type **A Non-automatic weighing instrument**
Type : PCS

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
4 March 2015

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).



OIML Member State
The Netherlands

Number R76/2006-NL1-14.41
Project number 13200647
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-08.07a dated 3 March 2008 that includes 37 pages;
- No. R76/1992-NL1-08.07b dated 1 March 2008 that includes 57 pages;
- No. NMI-11200059-01 dated 4 May 2011 that includes 15 pages;
- No. NMI-12200481-01 dated 3 December 2012 that includes 14 pages;
- No. NMI-11200059-02 dated 20 June 2014 that includes 28 pages;
- No. NMI-13200647-01 dated 11 February 2015 that includes 61 pages;
- No. NMI-13200647-02 dated 11 February 2015 that includes 14 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Maximum number of scale intervals	$n \leq 6000$ divisions for single interval instruments $n \leq 6000$ divisions per partial weighing range for multi-interval instruments
Maximum partial weighing ranges	2
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	$T \leq -\text{Max}$ for single interval instruments $T \leq -\text{Max}_1$ for multi interval instruments
Weighing range(s)	Single interval Multi-interval
Power supply voltage	230 V AC, 50/60 Hz 12 V battery (optional).
Application	Intended to be used for direct sales to the public
Software identification	See following table

EPOS software package	OS	Legally relevant files	Version	MD5 hash
ScalePos	Windows	Epelsa.dll	1.0.2.3	E3E57643A0178BE8C21F9CFF53BFB80D
TouchScale	Windows	Weight.dll	2	9669792B564F793471C962C02BD8D912
TouchScale	Linux	Libwgt.so	1	DCF11D09B220209E7F92EF48494D6DC2
OpenScale	Windows	DLLBPTC.dll	1.0.0.2	5CD1E726C4A66E268525B6718A0ECC66