

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-14.64  
Project number SO14204435  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	SATIS CO., LIMITED Flat B07 Floor 23, Hover Industrial Building No. 26-38 Kwai Cheong Road, N.T. Hong Kong
Identification of the certified type	An <b>Indicator</b> Type : STP-100L, STP-100S, STP-200C, STP-200E, STS-100L, STS-100S, STS-200C, STS-200E.
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
23 December 2014

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl)).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200302-01 dated 29 November 2013 that includes 51 pages;
- No. NMI-13200302-02 dated 29 November 2013 that includes 11 pages.

**Characteristics of the indicator:**

Accuracy class	(III)	
Maximum number of verification scale intervals	6000	
Load cell excitation voltage	5 V DC	
Minimum input voltage per verification scale interval	1 $\mu$ V	
Minimum load cell resistance	350 $\Omega$	
Maximum load cell resistance	1050 $\Omega$	
Fraction of the maximum permissible error	0,5	
Load cell connection	4-wire	6-wire (remote sensing)
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	connected directly without junction box	No special cable length
Weighing range(s)	Single interval Multiple range	
Maximum number of load platforms	1	
Temperature range	-10 °C / +40 °C	
Power supply voltage	230 V AC 50/60 Hz; AC/DC Adapter 9 V DC.	
Software identification	Version number: 1.xx for LCD mainboard, 2.xx for LED mainboard. Where xx represents the metrological non-relevant part of the software.	