

OIML Member State
The Netherlands

Number R76/2006-NL1-15.07 revision 1
Project number 13200543
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Hill-Rom, Inc. 1069 State Route 46 East Batesville, IN 47006 United States of America
Identification of the certified type	A Non-automatic weighing instrument Type : Compella
Characteristics	See next page


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
12 May 2015



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



OIML Member State
The Netherlands

Number R76/2006-NL1-15.07 revision 1
Project number 13200543
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-12200558-02 dated 17 July 2013 that includes 24 pages;
- No. NMI-13200543-01 revision 1 dated 7 May 2015 that includes 37 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max ≤ 500 kg
Verification scale interval	e ≥ 1 kg
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	n ≤ 500 divisions
Temperature range	+10 °C / +40 °C
Power supply voltage	230 V AC 50/60 Hz
Application	Intended to be used for weighing patients
Software identification	Version number: xxxx.01xx, where 01 represents the legally relevant software and xx is a number between 00 and 99

Revision History

This revision replaces the previous version(s).

Revision	Date	Change(s)
Initial	13 March 2015	-
1	12 May 2015	Update of the software version number to match the production version instead of the development version