

OIML Member State
The Netherlands

Number R76/2006-NL1-16.21
Project number 15200324
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Balanças Marques de José Pimenta Marques, Lda Parque Industrial de Celeirós (2a Fase) Apartado 2376 4701-905 Braga Portugal
Identification of the certified type	A Non-automatic weighing instrument Type : BMX0 BMX1
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class $\textcircled{\text{III}}$ or $\textcircled{\text{III}}$

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
4 March 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-15200324-01 dated 4 March 2016 that includes 40 pages;
- No. NMI-15200324-02 dated 4 March 2016 that includes 26 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	(III) or (III)	
Maximum capacity	$1,5 \text{ kg} \leq \text{Max} \leq 100 \text{ kg}$	
Verification scale interval	$e \geq 0,5 \text{ g}$	
Weighing range(s)	Single interval	Multi-interval
Maximum number of scale intervals	$n \leq 3000$ divisions (per partial weighing range)	
Maximum number of partial weighing ranges	2	
Tare	$T \leq -\text{Max}$	$T \leq -\text{Max}_1 - e_1$
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	100 –240 V AC 50/60 Hz	
Application	Intended to be used for direct sales to the public	
Software identification	BMX0	Version number: 2.xx xx= 00..99
	BMX1	Version number: 1.xx xx= 00..99

Software:

- The identification number will be displayed at start-up;
- The non-automatic weighing instrument has embedded software.