



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-16.22
Project number 15200560
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	InBody Co., Ltd. InBody Bldg., 54, Nonhyeon-ro 2-gil Gangnam-gu, Seoul 135-960 Korea
Identification of the certified type	A Non-automatic weighing instrument Type : InBody J30 / 230 / 270 / 370S / 470
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
18 April 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-16.22
Project number 15200560
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R76/1992-NL1-08.09 dated 19 June 2008 that includes 54 pages;
- No. NMI-11200383-01 dated 16 March 2012 that includes 30 pages;
- No. NMI-14200704-01 dated 30 January 2015 that includes 18 pages.
- No. NMI-15200506-01 dated 18 April 2016 that includes 22 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max \leq 200 kg
Verification scale interval	$e \geq 0,1$ kg
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 2000$ divisions
Temperature range	0 °C / +40 °C
Power supply voltage	12 V DC through 100 - 240 V AC 50/60 Hz adapter
Software identification (version numbers)	IB230-1.0 IB270-1.0 IB470-1.0 IBJ30-1.0 IB370S-1.0