

## OIML Certificate of Conformity

**OIML Member State** 

The Netherlands

Number R76/2006-NL1-16.25 Revision 1 Project number 1901563 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and

**Ohaus Corporation** 

Manufacturer

7 Campus Drive, Suite 310 Parsippany, NJ 07054

**United States of America** 

This Certificate does not bestow any form of legal international approval.

Identification of the

A Non-automatic weighing instrument

certified type

Type : RS-series

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

18 December 2017

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org







## OIML Certificate of Conformity

**OIML Member State** 

The Netherlands

Number R76/2006-NL1-16.25 Revision 1 Project number 1901563 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R76/1992-NL1-05.16 dated 3 June 2005 that includes 55 pages;
- No. R76/1992-NL1-06.38 dated 2 November 2006 that includes 44 pages;
- No. R76/1992-NL1-09.12 dated 6 April 2009 that includes 17 pages;
- No. NMi-12200063-01 dated 25 May 2012 that includes 26 pages.

## Characteristics of the non-automatic weighing instrument:

Load cell + + + + + + + + +	+ + + + AMI + + + + +	New AMI
Accuracy class	+ + + + + + + + + + + + + + + + + + +	
Maximum capacity	3 kg ≤ Max ≤ 30 kg	3 kg ≤ Max ≤ 30 kg
Verification scale interval	+ + + e ≥ 1 g + + + + +	+ + + e ≥ 1 g + + +
Weighing range(s)	Single interval	
Maximum number of scale intervals	n ≤ 3000 divisions	n ≤ 6000 divisions
Tare* * * * * * * * * * * * * * * * * * *	T ≤ -Max	T ≤ -Max
Temperature range	-10 °C / +40 °C	
Power supply voltage	100 - 240 V AC 50/60 Hz 12 V optional internal rechargeable battery	
Application	Intended to be used for direct sales to the public	
Software identification * * * *	179403 or 250797	

The software identification is displayed at start-up; The software is embedded.

## **Revision History**

This revision replaces the previous version.

Revision	Date	Change(s)
Initial	29 August 2016	<u> </u>
1 + + +	18 December 2017	Correction of the software identification (editorial) + + + +

5