

OIML Member State
The Netherlands

Number R76/2006-NL1-16.32
Project number 15200681
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo GmbH Im Langacher 44 8806 Greifensee Switzerland
Identification of the certified type	A Non-automatic weighing instrument Type : WX-series
Characteristics	See next page


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **I**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
19 April 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-08.45 dated 23 December 2008 that includes 39 pages;
- No. NMI-15200681-01 dated 19 April 2016 that includes 43 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	Ⓛ
Maximum capacity	$\text{Max} \leq 220 \text{ g}$ or the equivalent in ct
Verification scale interval	$e \geq 1 \text{ mg}$ or the equivalent in ct
Actual scale interval	For instruments where $n \geq 50000$: $d \leq e \leq 1000 d$; For instruments where $n \leq 50000$: $20 d \leq e \leq 1000 d$
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 220000$ divisions
Tare	$T \leq -\text{Max}$
Temperature range	$+10 \text{ }^\circ\text{C} / +30 \text{ }^\circ\text{C}$
Power supply voltage	$100 - 240 \text{ V AC } 50/60 \text{ Hz}$
Software identification	Version numbers: 1.xx (weighing cell software) (xx is a number between 00 and 99)