

OIML Member State
The Netherlands

Number R76/2006-NL1-16.60
Project number 15200386
Page 1 of 2

| | |
|--------------------------------------|---|
| Issuing authority | NMi Certin B.V. Person responsible: C. Oosterman |
| Applicant and Manufacturer | Arjo Huntleigh AB Hans Michelsengatan 10 211 20 Malmö Sweden |
| Identification of the certified type | A Non-automatic weighing instrument Type : Alenti, Miranti |
| Characteristics | See next page |

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
21 December 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



The conformity was established by the results of tests and examinations provided in the associated OIML Type evaluation report:

- No. NMI-15200386-03 Revision 1 dated 14 December 2016 that includes 52 pages;
- No. NMI-15200386-01 dated 14 December 2016 that includes 41 pages.

Characteristics of the non-automatic weighing instrument:

| | |
|--|---|
| Accuracy class | III |
| Maximum capacity | Max \leq 182 kg |
| Verification scale interval | $e \geq 0,1$ kg |
| Weighing ranges | Single interval Multiple range |
| Maximum number of scale intervals (one weighing range) | $n \leq 1600$ divisions |
| Maximum number of scale intervals (multiple range) | $n \leq 1200$ divisions |
| Maximum number of weighing ranges | 2 |
| Tare | $T \leq -(Max+9e)$ |
| Temperature range | +10 °C / +40 °C |
| Power supply voltage | 24 V DC by the battery pack of the lifter |
| Application | To be used only for weighing patients |
| Software identification | Version number: 0.xx y (xx = 00...99 / yy = A...Z) |