

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-16.61  
Project number 16200668  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	BISON Group Ltd Unit 2, 11 Wharf Street Dunedin 9016 New Zealand
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : Container Weighing System
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
5 December 2016



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-16.61  
Project number 16200668  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-15200539-01 dated 30 November 2016 that includes 42 pages;
- No. NMI-16200668-01 dated 30 November 2016 that includes 10 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$5000 \text{ kg} \leq \text{Max} \leq 175000 \text{ kg}$
Verification scale interval	$e \geq 50 \text{ kg}$
Weighing range(s)	Single interval
Maximum number of scale intervals	$n \leq 700$ divisions
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	6 V DC
Application	Intended to be used for weighing containers or other constructions fitted with four ISO 1161 corner castings
Software identification	Version number: 4.07.xx (xx is a number between 00 and 99)