

OIML Member State
The Netherlands

Number R76/2006-NL1-17.09
Revision 1
Project number 16200755
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Radwag Wagi Elektroniczne Bracka 28 26-600 Radom Poland
Identification of the certified type	A Non-automatic weighing instrument Type : MYA...4Y.. , UYA...4Y.. XA...4Y..
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **I**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
25 April 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-17.09
Revision 1
Project number 16200755
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-16200755-03 Revision 1 dated 25 April 2017 that includes 26 pages;
- No. NMI-16200755-04 dated 13 March 2017 that includes 41 pages.

Characteristics of the non-automatic weighing instrument:

	MYA...4Y.. , UYA...4Y..	XA...4Y..
Accuracy class	(I)	
Maximum capacity	Max ≤ 52 g, or Max ≤ 260 ct	Max ≤ 360 g, or Max ≤ 1800 ct
Verification scale interval	e ≥ 1 mg, or e ≥ 0,01 ct	
Actual scale interval	e = d, or e = 10 d, or e = 100 d, or e = 1000 d, or e = 10000 d	
Weighing range	Single interval	
Maximum number of scale intervals	n ≤ 52000 divisions	n ≤ 360000 divisions
Tare	T ≤ -Max	
Temperature range	+10 °C / +40 °C	
Power supply voltage	100 – 240 V AC 50/60 Hz	
Software identification	Version number for weighing firmware: 1.0.1	

Revision History

This revision replaces the previous version.

Revision	Date	Changes
Initial	13 March 2017	-
1	25 April 2017	For MYA...4Y.., UYA...4Y.. models, maximum capacity information is corrected from 110 ct to 260 ct. Test report NMI-16200755-03 is corrected to Revision 1. Reference to test report NMI-16200755-02 is removed.