



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.19
Project number SO16202631
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Ohaus Corporation 7 Campus Drive, Suite 310 Parsippany, NJ 07054 United States of America
Identification of the certified type	A Non-automatic weighing instrument Type : S51 S71
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
1 May 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.19
Project number SO16202631
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-13200324-01 dated 29 November 2013 that includes 59 pages;
- No. NMI-13200324-02 dated 15 October 2014 that includes 14 pages;
- No. NMI-16200297-01 dated 27 June 2016 that includes 8 pages;
- No. NMI-16200784-01 dated 21 February 2017 that includes 11 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing ranges	Single interval Multi-interval
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	$T \leq -\text{Max}$
Power supply voltage	100 – 240 V AC 50/60 Hz 12 V rechargeable battery 12 – 24 V road vehicle power supply
Application	Intended to be used for direct sales to the public
Software identification scale application	Version: "1.xxx.yyy" where "xxx" represents the legally non-relevant software part and "yyy" represents bug fixes.