

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.27  
Project number 1900904  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Kingship Weighing Machine Corp. 739, Renhua Rd. Dali Dist., Taichung 412 Taiwan R.O.C.
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : AEx, ASx, BEx, EQx, ESx, SEx, SSx, WDX, AMx, APx, ABPL, QEx, QSx (the x can be the letter P, W or C, and can be followed by X, L, L-M, or L-E)
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
16 June 2017



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-812089A dated 4 September 2009 that includes 46 pages;
- No. NMI-812089B dated 18 September 2009 that includes 12 pages;
- No. NMI-812089C dated 18 September 2009 that includes 12 pages;
- No. NMI-10200741-01 dated 21 February 2011 that includes 12 pages;
- No. NMI-11200515-01 dated 4 August 2011 that includes 11 pages;
- No. NMI-12200233-01 dated 14 June 2012 that includes 10 pages;
- No. NMI-14200138-01 dated 26 June 2014 that includes 26 pages;
- No. NMI-14200310-01 dated 7 July 2015 that includes 13 pages;
- No. R60/2000-NL1-07.18 dated 22 November 2007 that includes 40 pages;
- No. NMI-10200947-07 dated 24 December 2010 that includes 59 pages;
- No. NMI-10200947-08 dated 24 December 2010 that includes 59 pages;
- No. PTB 1.12-4039512-1 dated 22 January 2009 that includes 22 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 105 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing ranges	Single interval Multi-interval Multiple range
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Maximum number of scale intervals (multiple range)	$n \leq 3000$ divisions (per weighing range)
Maximum number of weighing ranges	2
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	110 or 230 V AC, 50/60 Hz Internal rechargeable battery 6 V DC Dry batteries, 6 V DC (4x D-cell 1,5 V DC each) External AC/DC adapter 230 V AC, 50 Hz
Application	Intended to be used for direct sales to the public
Software identification	Version number: 5.9 or 5.9-x- (x= 0...9)

The compatibility checking of modules of non-automatic weighing instruments according to Annex F of OIML R 76-1:2006 shall be fulfilled by the manufacturer at the time of putting into use.