

OIML Member State
The Netherlands

Number R76/2006-NL1-17.30
Project number 1901012
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	GRUPO EPELSA, S.L. c/. Punto Net, 3 Polígono Tecnológico TECNOALCALÁ E-28805 Alcalá de Henares, Madrid Spain
Identification of the certified type	A Non-automatic weighing instrument Type : PCSK
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
15 May 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-17.30
Project number 1901012
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200647-01 dated 11 February 2015 that includes 61 pages;
- No. NMI-13200647-02 dated 11 February 2015 that includes 14 pages;
- No. NMI-14200734-01 dated 7 July 2015 that includes 13 pages;
- No. NMI-14200734-02 dated 7 July 2015 that includes 13 pages;
- No. NMI-15200576-01 dated 27 October 2015 that includes 13 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	3 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 1 g
Maximum number of scale intervals	n ≤ 3000 divisions for single interval instruments n ≤ 3000 divisions per partial weighing range for multi-interval instruments
Maximum number of partial weighing ranges	2
Weighing ranges	Single interval Multi-interval
Tare	T ≤ -2/3 Max for instruments with one weighing range T ≤ -Max ₁ for multi-interval instruments
Temperature range	-10 °C / +40 °C
Power supply voltage	220 – 230 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public Intended to be used for the making-up of prepackages
Software identification	See following table

POS software package	OS	Legally relevant files	Version	MD5 hash
ScalePos	Windows	Epelsa.dll	1.0.2.3	E3E57643A0178BE8C21F9CFF53BFB80D
OpenScale	Windows	DLLBPTC.dll	1.0.0.2	5CD1E726C4A66E268525B6718A0ECC66
TouchScale	Windows	Libwgt.dll	1	9669792b564f793471c962c02bd8d912
TouchScale	Linux	Libwgt.so	1	dcf11d09b220209e7f92ef48494d6dc2
EBP	Windows	DLLEBP.DLL	1.0.0	D649387A90CDE835E3660F9C320D618F