



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.32  
Project number 1901013  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Applicant and Manufacturer GRUPO EPELSA, S.L.  
c/. Punto Net, 3  
Polígono Tecnológico TECNOALCALÁ  
E-28805 Alcalá de Henares  
Madrid  
Spain

Identification of the certified type **A Non-automatic weighing instrument**  
Type : UHRS

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
1 June 2017



C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.32  
Project number 1901013  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-08.07a dated 3 March 2008 that includes 37 pages;
- No. R76/1992-NL1-08.07b dated 1 March 2008 that includes 57 pages;
- No. NMI-12200481-01 dated 3 December 2012 that includes 14 pages;
- No. NMI-13200647-01 dated 11 February 2015 that includes 61 pages;
- No. NMI-13200647-02 dated 11 February 2015 that includes 14 pages;
- No. NMI-1901013-01 dated 16 May 2017 that includes 31 pages;
- No. NMI-10201103-01 dated 31 May 2017 that includes 8 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Maximum number of scale intervals	$n \leq 6000$ divisions for single interval $n \leq 3000$ divisions per partial weighing range for multi-interval
Maximum partial weighing ranges	2
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	$T \leq -\text{Max}$
Weighing range(s)	Single interval Multi-interval
Power supply voltage	120 - 230 V AC 50/60 Hz 12 V battery (optional)
Application	Intended to be used for direct sales to the public
Software identification	X-XXXXXXXX-XXX-1 (X may be a letter from A to Z or a number from 0 to 9, the last digit 1 is representing the metrological software part)