



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/1992-NL1-17.35
Project number 1901155
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	RAVAS Europe B.V. Toepadweg 7 5301 KA Zaltbommel The Netherlands
Identification of the certified type	A Non-automatic weighing instrument Type : iForksM / HyperforksM
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 1992 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
16 June 2017


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/1992-NL1-17.35
Project number 1901155
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R60/2000-NL-01.18 dated 5 September 2001 that includes 40 pages;
- No. R60/2000-NL-02.01 dated 3 January 2002 that includes 39 pages;
- No. 9200061 dated 17 May 2010 that includes 41 pages;
- No. 9200061B dated 9 July 2016 that includes 16 pages;
- No. NMI-16200696-01 dated 2 March 2017 that included 51 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	1000 kg ≤ Max ≤ 5000 kg
Verification scale interval	e ≥ 2 kg
Weighing range(s)	Single interval Multiple range
Maximum number of scale intervals (one weighing range)	n ≤ 1000 divisions
Maximum number of scale intervals (multiple range)	n ≤ 1000 divisions (per weighing range)
Maximum number of weighing ranges	2
Temperature range	-10 °C / +40 °C
Power supply voltage	Indicator: 6 or 12 V DC; Fork: 3,8 – 6,0 VDC.
Software identification	1.20C