

OIML Member State
The Netherlands

Number R76/2006-NL1-17.36
Project number 1901157
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome, Ohta-Ku, Tokyo 146-8580, Japan
Identification of the certified type	A Non-automatic weighing instrument Type : DIX-2001
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
14 July 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.36
Project number 1901157
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Indicator:

- No. NMI-16200772-01 dated 16 May 2017 that includes 53 pages.

Load cells:

- No. R60/2000-NL1-10.28A dated 10 January 2011 that includes 43 pages;
- No. R60/2000-NL1-10.28B dated 10 January 2011 that includes 43 pages;
- No. R60/2000-NL1-10.28C dated 10 January 2011 that includes 38 pages;
- No. NMI-10200371-01 dated 25 March 2011 that includes 35 pages;
- No. NMI-10200371-02 dated 25 March 2011 that includes 28 pages;
- No. NMI-10200372-01 dated 25 March 2011 that includes 35 pages;
- No. NMI-10200372-02 dated 25 March 2011 that includes 33 pages;
- No. NMI-10200372-03 dated 25 March 2011 that includes 28 pages;
- No. NMI-10200373-01 dated 2 May 2011 that includes 35 pages;
- No. NMI-10200373-03 dated 2 May 2011 that includes 28 pages;
- No. NMI-15200656-01 dated 10 November 2015 that includes 31 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III	III
Maximum capacity	3 kg ≤ Max ≤ 1500 kg	
Verification scale interval	e ≥ 1 g	
Weighing range(s)	Single interval	
Maximum number of scale intervals (one weighing range)	n ≤ 6000 divisions	n ≤ 1000 divisions
Power supply voltage	100 – 240 V AC 50/60 Hz 24 V DC rechargeable battery	
Software identification	Version number: 5.30	

The compatibility checking of modules of non-automatic weighing instruments according to Annex F of OIML R 76-1:2006 shall be fulfilled by the manufacturer at the time of putting into use.