

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.43  
Project number 1901086  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Ohaus Corporation 7 Campus Drive, Suite 310 PARSIPPANY NJ, 07054 United States of America
Identification of the certified type	A <b>Non-automatic weighing instrument</b> Type : Scout SJX series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
11 July 2017



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.43  
Project number 1901086  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-1520057-01 dated 1 September 2016 that includes 44 pages;
- No. NMI-1520057-02 dated 1 September 2016 that includes 16 pages;
- No. NMI-1520057-03 dated 1 September 2016 that includes 32 pages;
- No. NMI-1520057-04 dated 1 September 2016 that includes 24 pages;
- No. NMI-1901086-01 dated 4 July 2017 that includes 20 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	II
Maximum capacity	$64 \text{ g} \leq \text{Max} \leq 6200 \text{ g}$ , or $320 \text{ ct} \leq \text{Max} \leq 31000 \text{ ct}$
Verification scale interval	$e \geq 0,01 \text{ g}$ , or $e \geq 0,05 \text{ ct}$
Actual scale interval	$e = d$ , or $e = 10 d$
Weighing range(s)	Single interval
Maximum number of scale intervals	$n \leq 15000$ divisions
Tare	$T \leq -\text{Max}$
Temperature range	$+10 \text{ }^\circ\text{C} / +30 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz, or 6 V DC (battery)
Software identification	Version number: Sr 1.xx (xx is 00 to 99)