



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.47  
Project number 1901421  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Shanghai Teraoka Electronic Co., Ltd. Tinglin Industry Developmental Zone Jin Shan County Shanghai 201505 P.R. of China
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : RM-5800
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
25 August 2017



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/2006-NL1-10.06 dated 22 March 2010 that includes 59 pages;
- No. NMI-12200263-01 dated 28 September 2012 that includes 23 pages;
- No. NMI-12200684-01 dated 6 December 2012 that includes 11 pages;
- No. NMI-13200468-01 dated 21 October 2013 that includes 21 pages;
- No. NMI-14200357-01 dated 10 October 2014 that includes 37 pages;
- No. NMI-14200490-01 dated 10 October 2014 that includes 17 pages;
- No. NMI-15200739-01 dated 24 February 2016 that includes 35 pages;
- No. NMI-1900734-01 dated 20 April 2017 that includes 42 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class		(III)			
Maximum capacity		$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$			
Verification scale interval		$e \geq 1 \text{ g}$			
Weighing ranges		Single interval and / or Multi-interval			
Maximum number of scale intervals		$n \leq 6000$ divisions (single interval, with A/D board STB-2236), or $n \leq 3000$ divisions (per partial weighing range)			
Maximum partial weighing ranges		2			
Tare		$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments			
Temperature range		$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$			
Power supply voltage		100 – 240 V AC 50/60 Hz			
Application		Integrated POS software Intended to be used for direct sales to the public			
Software identification		Filenames	Version	Checksum	Remarks
Scale software		libbwm.jar	1.1.0.release_build:59 1.1.0.release_build:51 1 (Maxlabel software)	00085704 00077148 14015C4F	-
		libscale.jar libprice.jar	1.1.0.release_build:26 1.1.0.release_build:5	00064855	one checksum calculated over two files
POS software	Windows	libscale.dll libprice.dll	1 1	00459169 or 41047A60	
	Linux	libscale.so libprice.so	1 1	00747997 or 169330BA	
		Rm5800drv.so	1.0b145	8293B120200D92BF5F8 6EC2A4D099F9E	-