

OIML Member State
The Netherlands

Number R76/2006-NL1-17.50
Project number 1901207
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Arjo Huntleigh AB Hans Michelsengatan 10 211 20 Malmö Sweden
Identification of the certified type	A Non-automatic weighing instrument Type : Maxi Move
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
3 October 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.50
Project number 1901207
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report:

- No. NMI-1901207-01 dated 2 October 2017 that includes 51 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max \leq 227 kg
Verification scale interval	$e \geq 0,1$ kg
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 2270$
Tare	$T \leq -\text{Max}$
Temperature range	+10 °C / +40 °C
Power supply voltage	24 V DC by the battery pack of the lifter
Application	To be used only for weighing patients
Software identification	Version number: 3.xx (y, x = 0... 99)